

JUN 30 2014

## 5. 510(k) Summary

This summary of 510(k) safety and effectiveness information is being submitted in accordance with requirements of 21 CFR Part 807.92.

### 5.1 Company and Correspondent making the submission

Name	IntroMedic Co., Ltd.
Address	Suite 1104, E&C Venture Dream Tower 6-Cha, 197-28 Guro-Dong, Guro-Gu, Seoul, 152-719 Korea
Telephone	+82-2-801-9300
Fax	+82-2-801-9330
Contact	JinYoung Lee
Internet	<a href="http://www.intromedic.com">http://www.intromedic.com</a>

### 5.2 Device

Proprietary Name	MiroCam® Capsule Endoscope System
Common Name	Capsule Imaging System
Classification Name	21 CFR 876.1300 (Product Code NEZ)

### 5.3 Predicate Device

Manufacturer	IntroMedic Co., Ltd.
Device	MiroCam® Capsule Endoscope System
510(k) Number	K123428

Manufacturer	Given Imaging Ltd.
Device	Given PillCam® Platform with PillCam® SB Capsules with PillCam® SensorBelt
510(k) Number	K103088

#### 5.4 Description

(※ The device description is exactly the same as the predicate device (K123428).)

The MiroCam® Capsule Endoscope System is comprised of the following core components:

- MiroCam® Capsule: MC1000-B
- MiroCam® Receiver: MR1100
- Data Cable / Data Belt (Optional): MR1000-D / MR1000-D(M) or MR1000-D(L)
- MiroView™ Software: MiroView™ U, MiroView™ RTV, MiroView™ RTV-i

The general usage workflow of the MiroCam® system is as follows, the MiroCam® capsule captures images of the GI tract, which are sent via Human Body Communication to sensor pads which are affixed to the patient body. The sensor pads are connected to the receiver by the data cables. The image data is stored on the receiver for the duration of the patient procedure. After removing the receiver set from the patient body, the receiver is connected via USB to the Commercial PC and the image data is uploaded. Following upload, the physician (Gastroenterologist) reviews the patient image data for suspected abnormalities of the small bowel. The key system components are explained in detail below.

##### MiroCam® Capsule

Functionally, the MiroCam® capsule captures imaging via a CMOS imaging sensor for at least 12 hours at the rate of 3 images per second. Six white LEDs flash in concert with the imaging sensor. A sensor PCB links the imaging sensor to the two gold electrodes of the capsule, from which a weak current is emitted containing the image data via Human Body Communication technology. The weak current is then passed through the bodily tissue and fluid to be picked up by the sensor pads. Human Body Communications is uni-directional from the capsule to the sensor pads.

Physically, the capsule is 24.5mm in length, and 10.8mm in diameter. The exterior of the

capsule is composed of biocompatible materials, capable of withstanding potential bite forces and exposure to fluids in the GI tract.

#### MiroCam® Receiver

The image data from the capsule is saved to the MiroCam® Receiver via the sensor pads and data cables. The sensor pads are standard ECG type sensor pads (3M Red Dot), which are affixed to the patient's abdomen according to the location guide. The 9 sensor pads are attached to the 9 leads of the data cable, and the data cable is subsequently attached to the Receiver. The Receiver sits in a shoulder pouch, which is put on the patient for the duration of the procedure. A waist pouch is used to organize the data cables. The Receiver can store images for up to 12 hours. The three indicators on the Receiver are for battery status, initialization status of the receiver, and status of signal receipt from the capsule.

The dimensions of the receiver are 140mm in height, 85mm in width, and 40mm in length. The units weight 350 grams with the battery, which is data cable. The battery is lithium ion, and is unique to the Receiver. The battery is rechargeable, and can be charged by placing in the battery charger provided with the Receiver. The receiver uses a storage flash for storing the images.

The MiroCam® Receiver has RTV (Real-Time View) function to view a capsule in the body in real time. There are the wire communication method to connect to a Commercial PC by USB cable and the wireless communication method to connect to i-Phone, i-Pad or i-Pod touch by Wifi for RTV.

#### MiroView™ U

The MiroView™ U helps the user to upload the image data from the Receiver and review the image data with displaying.

MiroView™ U will be deployed as three types of software: MiroView™ Server, MiroView™ Operator and MiroView™ Client

Server PC includes two types of SW: MiroView™ Server and Operator

MiroView™ Server and MiroView™ Operator will be deployed on a Server PC. MiroView™ Client SW will be deployed on reviewer's PCs. 10 Client SWs will be able to connect to one

Server PC via Network (LAN) at a time.

MiroView™ Server has functionalities: Data Storage, PACS connectivity

MiroView™ Operator has functionalities: Receiver Management, Case Management and Account Management

MiroView™ Client has functionalities: List Mode (Case Data Management), Review Mode (Video Display, Image Capture, Annotation Box, etc.), Report Mode (User Comment, Auto Disease Name Check, etc.), Export Mode (JPG, AVI, MiroView™ file format)

MiroView™ U provides some kinds of Modes and Views to help the user review patient video data quickly and conveniently; however, IntroMedic Co., Ltd. recommends that the doctor review the patient video in the Normal Mode / Single View to prevent from a wrong diagnosis.

The features are described below:

- Multi-Data Path (MiroView™ Server)
- Customizing Print Dialog (MiroView™ Client)
- Data Multi-Language (MiroView™ Client)
- ODF Report (MiroView™ Client)
- Data Bag (MiroView™ Client)
- Finding Setting (Custom Finding) (MiroView™ Client)

## **5.5 Indications for Use**

The MiroCam® Capsule Endoscope System is intended for visualization of the small bowel mucosa. It may be used as a tool in the detection of abnormalities of the small bowel in adults and children from two years of age.

The Suspected GI Bleeding Indicator (SGIB) is intended to mark frames of the video suspected of containing blood or red areas.

## **5.6 Technological Characteristics**

The technology characteristics are exactly the same as the predicate device (K123428).

### **5.7 Performance Testing**

The performance testing are exactly the same as the predicate device (K123428).

### **5.8 Conclusion**

Based on the 510k submission data IntroMedic Co., Ltd. believes that MiroCam® Capsule Endoscope System and the predicate devices selected are substantially equivalent and do not change the fundamental scientific technology and indications for use of the market-cleared devices.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

June 30, 2014

Intromedic Co., Ltd.  
% Jinyoung Lee  
General Manager  
Suite 1104 E&C Venture Dream Tower 6-cha  
197-28 Guro-dong, Guro-gu  
Seoul 152-719  
Korea

Re: K140751  
Trade/Device Name: Mirocam® Capsule Endoscope System  
Regulation Number: 21 CFR§ 876.1300  
Regulation Name: Ingestible telemetric gastrointestinal capsule imaging system  
Regulatory Class: II  
Product Code: NEZ  
Dated: May 2, 2014  
Received: May 5, 2014

Dear Jinyoung Lee,

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies.

You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Benjamin R. Fisher -S

Benjamin R. Fisher, Ph.D.  
Director  
Division of Reproductive, Gastro-Renal,  
and Urological Devices  
Office of Device Evaluation  
Center for Devices and Radiological Health

Enclosure

#### **4. Indications for Use**

**510(k) Number (if known):** K140751

**Device Name:** MiroCam® Capsule Endoscope System

**Indications for Use:**

The MiroCam® Capsule Endoscope System is intended for visualization of the small bowel mucosa. It may be used as a tool in the detection of abnormalities of the small bowel in adults and children from two years of age.

The suspected GI Bleeding Indicator (SGIB) is intended to mark frames of the video suspected of containing blood or red areas.

Prescription Use \_\_\_\_\_ AND/OR Over-The-Counter Use \_\_\_\_\_  
(Part 21 CFR 801 Subpart D) (Part 21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

---

Concurrence of CDRH, Office of Device Evaluation (ODE)

**Benjamin R. Fisher -S**  
**2014.06.30 16:32:18 -04'00'**